APEXPILOT® 61

Endo Motor

Adopt real-time feedback technology and dynamic torque control, realizing automatic reverse, effectively preventing fracture.





Imported Motor

15.5:1 reduction ratio endows strong power and efficient cutting. Durable planet gear gearbox prolongs service life of device.



Automatic Torque Calibrating

Choose torque calibrating mode to automatically calibrate torque, reducing the torque fluctuation caused by contra-angle replacement or sterilization under high temperature and high pressure.



Wireless Battery Charging

Contactless charging without contact point outside well avoids poor contact. Build-in battery, charging by putting handpiece into charging base.



Nimble Contra-angle

Contra-angle can be autoclaved to avoid crossinfection. Adjustable angle make it available for different teeth positions. Small and exquisite head provides unobstructed views.

Key Specifications

 Speed Range
 100-1000RPM

 Torque Range
 0.6Ncm-5.0Ncm

 Lithium Battery
 3.6V /750mAh

Power Input ~100V-240V 50Hz/60Hz 0.4A Max

Power Output DC5V/1A

Environment Temperature $+5^{\circ}\text{C} \sim +40^{\circ}\text{C}$ Relative Humidity $30\% \sim 75\%$

Atmospheric Pressure 70kPa ~ 106kPa

Warranty 12 Months

Order Information

Code: AL2050 Model: ApexPilot G1





Beyes Dental Canada Inc. 23-595 Middlefield Road Toronto, Ontario, M1B 3M5 Canada

Tel: 1-855-603-1888 Email: info@beyes.ca Web: www.beyes.ca



Lotus NL B.V.
Koningin Julianaplein 10, le Verd, 2595AA, The Hague,
Netherlands

Netherlands
Tel: +31645171879 (English)
+31626669008 (Dutch)
Email: info@beyes.ca

Your Dealer for Beyes Products







Beyes Canada Inc 2021. All Rights Reserved.
Specifications are subject to change without notice.
All brands are holders of their respective trademarks.
All Beyes products are manufactured in strict compliance with Health Canada regulations ensuring quality.
Manufacturing facility is ISO 13485 certified under Canadian Medical Devices Conformity Assessment System (CMDCAS).