



Instructions For Use Contra Angle -Plus

Table of Contents

- Before You Begin
- Safety Information
- Technical Specifications
- Handpiece and Bur Insertion
- Handpiece and Bur Removal
- Lubrication
- Sterilization
- Warranty



WARNING!
(Risk of injury)



ATTENTION!
(To prevent damage)



Thermal washer
disinfectable



Sterilizable
up to the stated
temperature

Before You Begin

Please read this Operator's Manual prior to using your Slow Speed Attachments. The use of these products is intended for use by trained dental personnel. We recommend that you retain this manual for future reference. This manual includes several attachment models so please refer to the Technical Specifications table to insure the model of your attachment is included.

Safety Information

- Slow speed handpieces and attachments should be sterilized after each patient to prevent any transmission of infectious organisms.
- Suitable hygienic measures must be taken to prevent cross contamination between patients, operators and other personnel.
- Only operate this slow speed handpiece when the attachments and burs are properly seated. Do not attempt to extend the length of a bur by improper positioning in the chuck.

Technical Specifications

Model	M29-MH01	M29L-MH01
Fiber-Optic	No	Yes
Bur Type	Push Button, CA Burs	Push Button, CA Burs
Water Spray	Internal Spray	Internal Spray

Mounting attachments to the motor

Slide the attachment onto the nose of the slow speed motor until it clicks into place. Check for secure connection. To remove an attachment, simply pull the attachment away from the motor. Fig 1




Fig 1

Bur Insertion

FG Bur
Insert the bur into the head. Push the button and insert the bur fully. Fig 2




Fig 2

Latch Bur
Insert the latch bur into the head. Slightly rotate the bur until it engages with the latch mechanism. Continue to push it inward until it clicks. Fig 3




Fig 3

1. Swing the Latch Plate to open the chuck. Fig 4
2. Insert the Latch bur.
3. Swing the Latch Plate back to closed position.




Fig 4

Bur Removal

Push the cap on the back of the head and pull bur to remove.




Fig 6

Swing the Latch Plate to open position and pull bur to remove. Fig 7




Fig 7

Lubrication

Lubricate with Beyes Spray Lubricant after each use and/or before autoclaving. Fig 8

- Mount the E-type spray nozzle into the spray port.
- Shake the can 3-4 times to well mix lubricant and propellant.
- Insert the E-type spray nozzle in the rear of the handpiece and spray for
- Approximately 2-3 seconds until the oil comes out of the handpiece



Fig 8

Caution

- Be sure to hold the handpiece firmly to prevent the handpiece from slipping out of hand by the spray pressure when lubricating.
- Supply lubricant until it comes out of the handpiece head (for approx. 2 seconds).
- Hold the Beyes Spray can upright.

Sterilization

Sterilization of instruments by applying a fractionated pre-vacuum steam sterilization process (according to EN285/EN 13060/EN ISO 17665) under consideration of the respective country requirements.



Recommended sterilization temperature: 135°C(±2°C), holding time: 5 min.

Minimum drying time after sterilization: 8 minutes.

Flash sterilization is not allowed on lumen instruments!



CAUTION:

Use steam sterilization exclusively for complete sterility. Never dry heat sterilize, ultrasonically clean, or expose handpiece or swivel to chemical disinfection or cold sterilization by:

- Immersion techniques or surgical milk
- Surface wipes, Hand soaps, or Detergents
- Surface sprays
- Sterilization bags containing disinfectants

Limited Warranty

Beyes Dental Canada hereby warrants against defect in material and workmanship that for a period of 12 months from the date of purchase, this product will perform satisfactorily under normal use and service. No warranty claim can be inferred here, as wear may occur earlier or later than indicated above depending on use, frequency of sterilization and frequency of maintenance. Beyes shall under no circumstances be liable for incidental or consequential damages or damage due to improper use or maintenance.

This Warranty is subject to the following conditions

Beyes will, without charge repair or replace (at Beyes sole discretion) any defective part/s covered in this warranty. If the conditions are met, the defective item must be returned to an authorized Beyes service center or distributor. Accompanied by the original invoice or packing list. This warranty is in lieu of all other warranties, expressed or implied.

* This device is not field repairable

* All brands are holders of their respective trademarks.



Federal law restricts this device to sale by or on the order of a dentist, physician, or any other practitioner licensed by the law of the states in which he or she practices to use or order the use of this device.

Beyes Dental Canada Inc.
23-595 Middlefield Rd
Scarborough, ON M1V3S2
Canada
Tel: 1-416-281-1888
Email: info@beyes.ca
Web: www.beyes.ca



Lotus NL B.V.
Koningin Julianaplein 10, le Verd,
2595AA, The Hague,
Netherlands



0197

Document #: ENI073R1
10.11.24